

Light Refractive Tomography for Noninvasive Ultrasound Measurements in Various Media

Refraktometrische Tomographie
zur berührungslosen Messung von
Ultraschall in unterschiedlichen Medien

Der Technischen Fakultät
der Friedrich-Alexander-Universität
Erlangen-Nürnberg
zur
Erlangung des Doktorgrades Dr.-Ing.

vorgelegt von

Lizhuo Chen

aus Zhejiang

Als Dissertation genehmigt von
der Technischen Fakultät der
Friedrich-Alexander-Universität Erlangen-Nürnberg

Tag der mündlichen Prüfung:	16.12.2013
Vorsitzende des Promotionsorgans:	Prof. Dr.-Ing. Marion Merklein
Gutachter:	Prof. Dr.-Ing. Reinhard Lerch
	Prof. Dr. techn. Bernhard G. Zagar

Berichte aus der Akustik

Lizhuo Chen

**Light Refractive Tomography for Noninvasive
Ultrasound Measurements in Various Media**

D 29 (Diss. Universität Erlangen-Nürnberg)

Shaker Verlag
Aachen 2014

Bibliographic information published by the Deutsche Nationalbibliothek

The Deutsche Nationalbibliothek lists this publication in the Deutsche Nationalbibliografie; detailed bibliographic data are available in the Internet at <http://dnb.d-nb.de>.

Zugl.: Erlangen-Nürnberg, Univ., Diss., 2013

Copyright Shaker Verlag 2014

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording or otherwise, without the prior permission of the publishers.

Printed in Germany.

ISBN 978-3-8440-2528-6

ISSN 1611-1303

Shaker Verlag GmbH • P.O. BOX 101818 • D-52018 Aachen

Phone: 0049/2407/9596-0 • Telefax: 0049/2407/9596-9

Internet: www.shaker.de • e-mail: info@shaker.de

Acknowledgements

I worked on this dissertation during my three-year employment in the Chair of Sensor Technology (LSE) at the University of Erlangen-Nuremberg, Germany. Here I would like to express my sincere gratitude to all those who provided inspiration and counsel for my thesis in these years.

First of all, I want to thank my supervisor, Prof. Dr. Reinhard Lerch, for his scholarly advice and valuable guidance. Especially appreciated is his encouragement, which kept me on course when in doubt. I would also like to thank for the comfortable working environment he established at our department.

I am further grateful to Prof. Dr. Bernhard G. Zagar who agreed to be the second assessor of this thesis without hesitation, and thank for his constructive comments and excellent advice.

Dr. Stefan J. Rupitsch has been my mentor, colleague and friend. His guidance made the work of the last three years an interesting and rewarding journey. His help was not restricted to the academic sphere. As a matter of fact, his advice on communication skills was particularly appreciated.

I greatly appreciate the help of Manfred Pelz for his support in constructing components for measurements. Thanks are due for his excellent work, which made the control of error sources simple. Because of this, accurate measurements of LRT have been successfully accomplished.

Moreover, I would like to thank Jens Grabinger, who provided consistent support in the field of numerical simulations, as well as Felix Wolf and Michael Löffler, who had come up with the idea of minimizing the placement errors with the aid of a laser beam. I also want to thank Martin Meiler, who offered me the opportunity of applying LRT to the investigation of a high-intensity ultrasound device, and his help in obtaining the permission of publishing the corresponding measurement results.

In the end, I gratefully acknowledge funding of the Erlangen Graduate School in Advanced Optical Technologies (SAOT) by the German Research Foundation (DFG) in the framework of the German excellence initiative.

Erlangen, December 2013

Lizhuo Chen

Contents

Notations and Abbreviations	vii
Abstract	ix
Kurzfassung	xi
1 Introduction	1
1.1 Motivation	1
1.2 State of the Art	2
1.3 Overview of the Work	4
1.4 Outline of the Thesis	7
2 Measurement Fundamentals	9
2.1 Measurement Principle	9
2.1.1 Sound Pressure in Fluids	10
2.1.2 Dilatation in Solids	10
2.2 Laser Doppler Vibrometer	11
2.3 Tomographic Imaging	13
2.3.1 Fourier Slice Theorem	13
2.3.2 Reconstruction for Parallel Projections	15
2.4 Interim Summary	16
3 Design of LRT Measurements	17
3.1 Experimental Setup	17
3.2 Choice of Measurement Parameters	19
3.2.1 Theoretical Determination of Measurement Parameters	19
3.2.2 Optimization of Measurement Parameters	21
3.3 Control of Measurement Errors	25
3.3.1 Placement Errors	25
3.3.2 Optical Errors	29
3.4 Measurable Frequency Range	32
3.5 Interim Summary	36
4 Ultrasound Pressure Fields in Water	37
4.1 Piston-type Transducer	37
4.2 Cylindrically Focused Transducer	41
4.3 Acceleration of Measurement Process	42
4.3.1 Reduction of the Number of Projections	43
4.3.2 LRT Measurements under the Axisymmetric Assumption	44
4.3.3 Three-dimensional Reconstruction of a Pressure Field	48

4.4	Reconstruction of a Disturbed Pressure Field	50
4.4.1	Numerical Simulation	50
4.4.2	Experiment and Results	50
4.5	Interim Summary	57
5	Ultrasound Pressure Fields in Air	59
5.1	Experimental Setup	59
5.2	Calculation of Piezo-optic Coefficient in Air	60
5.3	Experimental Results	62
5.4	Interim Summary	64
6	Ultrasound Fields in PMMA	67
6.1	Normal Stress in Isotropic Solids	67
6.2	Experimental Results	71
6.2.1	Ultrasound Fields Arising from Different Transducers	71
6.2.2	Ultrasound Fields throughout Water and PMMA	74
6.3	Interim Summary	80
7	High-intensity Ultrasound in Water	83
7.1	Experimental Setup	83
7.2	Measurement Results	84
7.2.1	Frequency Distribution	84
7.2.2	Repeatability	89
7.2.3	Reconstruction and Visualization	90
7.3	Interim Summary	96
8	Conclusion and Outlook	97
8.1	Conclusion	97
8.2	Outlook	100
	Appendix	103
	Bibliography	107
	List of Figures	113
	List of Tables	115